

CARETAKERSENTRY

with Two-way voice pendant

DEALER SET-UP & PROGRAMMING INSTRUCTIONS Model # 41911



THESE INSTRUCTIONS ARE FOR THE DEALER. REFER TO THE "QUICK START GUIDE" FOR BASIC INSTALLATION.

These instructions are for the Dealer to set-up and program the CaretakerSentry System to operate with the Central Station of your choice.

The CaretakerSentry product (the "System") uses the industry standard "Contact ID" and "4+2" protocols.

Voice prompts provide easy set-up and testing.

Basic set-up requires you to:

- Install the 4 AA rechargeable Ni-MH batteries in the Base unit.
- Plug the base unit into AC power (not a switched outlet)
- Connect to a wired telephone connection – Line In

Basic programming requires you to enter:

- The central station telephone number(s).
- The 4 digit account code of your customer.

There are several special options you can set for additional features - these include:

- Changing the 4 digit factory passcode to a passcode of your choice – to prevent unauthorized reprogramming
- You should enter two telephone numbers - the primary number should be the alarm receiver's toll free number and the secondary number should be the alarm receiver's local long distance number, in case the first number is busy or unreachable.
- You can program the System to send Restore Codes.
- You can program the System to send a Periodic Test Signal Every Day, Every Week or every 30 days.
- The System can learn up to a total of 9 Pendants, with up to 5 Two-Way Pendants and/or EWCs plus 4 Basic Pendants for a total of 9.

NOTE: At any time in the future, you can remotely change all of these setting over a telephone line using any touch-tone phone.

NOTICE - Please read all of the Limitations of Liability, Disclaimer and Limited Warranty that are in the User Quick Install Guide.

TABLE OF CONTENTS

System Components.....	4
Limitation of Liability.....	5
Installing the Batteries.....	6
Connecting the Base Unit.....	6
Power Connection.....	6
Telephone Connections.....	7
Communication Protocol Options.....	7
Line Seizure.....	7
Dealer Programming.....	8
Key to LED Status.....	8
Programming Central Station, Account Codes and Caretaker numbers...9	9
Changing a 4 Digit Passcode.....	9
Setting Programming Options.....	10
Remotely Programming a System.....	13
Adding Additional Pendants to the System.....	14
Erasing All Previously Learned Pendants.....	15
Basic Set-up and Test.....	15
Lanyard or Wrist-Strap.....	15
System Check.....	15
Operating Range.....	16
Testing.....	16
Out of Service Storage or Transportation.....	16
Troubleshooting Guide.....	17
Central Station Items of Interest.....	18
FCC Information.....	19
Important Safety Instructions.....	20
LogicMark Disclaimers and Limited Warranty.....	20
Replacement Parts / Optional Accessories.....	22
Product Specifications.....	23

System Components:

The items below are included with your System .



Pendant Part #41915



Base Unit Part #40914



Rechargeable Li-Ion Batteries - 2 Pieces
Part #35917



Rechargeable AA Batteries - 4 Pieces
Part #35918



AC Adapter



Phone Cord



Lanyard



Wrist strap and Beltclip
(Holster screws onto clip for beltclip
or strap for wrist strap)

Limitation of Liability

This page is a summary of the Disclaimer and Limited Warranty disclosed in full at the end of this instruction manual. Read these carefully.

The purchaser agrees by using this CaretakerAlert (the "System") to the terms and conditions below and in the Disclaimer and Limited Warranty. The purchaser also agrees to read and follow all instructions and warnings on the product and contained within these Installation and Operation Instructions.

It will be the sole responsibility of the purchaser and user to assure that this System is installed and programmed properly, and that the unit is used and maintained correctly. This includes, but is not limited to, periodic use to assure that the System and batteries, are in proper working order, that the unit is located in an appropriate location in the home, that the electrical outlet is supplying power, and that the user has been educated as to the operation and functionality of the product as a whole.

The System's equipment is not designed or guaranteed to prevent any loss or injury. This Limited Warranty and Disclaimer of Liability discussed in full at the end of this instruction manual constitutes the terms of sale and use of the System (and accessories) equipment, and if, notwithstanding these terms of sale and use of the product, there should arise any liability on the part of LogicMark LLC as a result of any cause whatsoever, regardless of whether or not such loss, damage, or personal injury was caused by or contributed to by LogicMark's negligence to any degree or failure to perform any obligation or strict products liability, such liability will be limited to an amount paid by the Purchaser for the product.

Further, LogicMark has no obligation to assure that calls are made, received or responded to, nor is LogicMark responsible for acts, or consequences of the acts, of those responding.

LogicMark provides no monitoring service for this product. It is up to the persons at the numbers called to respond in an appropriate manner.

Installing the Batteries:

Base - larger batteries: The Base unit requires 4 rechargeable 2400 mAH NiMH batteries (included). They will provide you with up 24 hours of battery back- up protection in the event of a power failure. They will not be damaged by continually charging in the base unit. The ratings on these can be 2400 mAh or higher.

When you first get your unit - these batteries will probably be low, or even fully discharged - and will need to be charged overnight before being able to provide you the 24 hour battery back-up protection.

Remove the cover on the bottom of the Base unit. Install the 4 batteries as shown in the plastic engraving in the battery compartment area. A few seconds after you install these, you might hear "Running on battery power" and RED Power LED will flash - when the unit is operating on the backup batteries.

(If you do not hear this announcement and the Power LED is on solid, the unit is already plugged in to AC power).



Master ON/OFF Switch:

CaretakerSentry has a Master ON/OFF switch on the bottom of the base unit – just right of the rechargeable batteries compartment on the base unit. This switch controls all power – back up batteries and power from the AC adapter. This switch should be ON at all times while the unit is in use. This switch should only be turned OFF if the CaretakerSentry unit is taken out of service – during transportation or no longer needed.

Connecting the Base Unit:

Power Connection:

Simply plug the AC Adapter into the nearest electrical outlet. Plug the other end of the AC power adapter into the Base unit. You need to push the plug into the opening in the Base unit and twist 90° clockwise to lock in place. To remove in the future, simply rotate 90° counter-clockwise and pull outward.



Caution: Do not plug the AC Adapter into an outlet which is controlled by a switch. The switch could accidentally be turned off, thus rendering the Base unit inoperable after approximately 24 hour when the back-up batteries become discharged.

Telephone Connection:

A standard telephone line is required to use your System. Your phone will work with your existing touch-tone phone line and existing service. It works the same way as any cordless phone.

NOTE: Touch Tone service is required.

Simply plug one end of the telephone cord into a telephone jack, and the other end into the "**Line In**" jack at the back of the Base unit. Note - you will feel a 'click' when the cords are firmly seated in the wall and the Base unit jack receptacles.

If you need to plug a conventional telephone into same outlet used by the Base unit, simply plug the telephone into the spare jack labeled "To House Phone."

Communication Protocol Options:

The CaretakerSentry unit supports 2 different alarm communication protocols – **Contact ID** and **4+2** - to help the installer deal with any type of phone System that the customer may be using:

Contact ID alarm communication protocol is not recommended for use with VoIP (**Voice over Internet Protocol**) internet phone Systems such as Skype, MagicJack, Basic Talk, Vonage, Comcast, Fox, etc. Presently the reliability of such services is not adequate for use with Contact ID in an emergency situation. **Check with your central station for recommendations when your customer is using a VoIP telephone service.**

4+2 is a "pulsed" alarm communication format that is much more reliable communication protocol with VoIP phone Systems – but not all Central Stations support the use of 4+2 communication protocol. Please check with your Central Station for recommendations before using 4+2 on a VoIP phone System.

How to Handle VoIP and other Internet phones: Check with your central station for recommendations when your customer is using a VoIP telephone service.

Line Seizure:

The phone plugged into the "To House Phone" jack will have the line seizure feature. This means that whenever the Base unit chooses to dial out during an emergency - it will seize the line from the phone that is plugged into this "To House Phone" jack if it is in use - and disconnect it - so that the Base unit can dial out. This is known as 'Local line seizure'.

Notice: If you choose to have 'Whole house line seizure', you need to run all phones through this "To House Phone" jack. This can only be done with professional installations.

Caution - If the Line Seizure feature is used and hooked-up, verification of Line Seize capability should be made immediately after installation, and periodically thereafter, in order to ensure that this equipment can initiate a call even when other equipment (telephone, answering System, computer modem, etc.) connected to the same line is in use.



Dealer Programming Instructions:

The System has several features to help with the programming. The System uses voice prompts and LED lights to guide you through the process. They give you a good audible confirmation of the progress as you are programming. Second are the LED lights on the base unit and the Pendant.

Key to LED Status - Base Unit:

YELLOW **BACK-UP BATTERY STATUS** LED functions:

- Off: **Running on Battery Power**
- Solid: Normal Operation – Charging Batteries
- Flashing: **Low or Defective Batteries**

BLUE **PHONE CONNECTION** LED function:

- Off: Normal operation
- Solid: Dialing/Talking
- Flashing: **No Phone Connection Detected**

GREEN **INFO** LED functions:

- Off: Standby
- Solid: Dialing/Talking or Learn Mode
- Flashing: Pendant Learning or **Problem with Pendant (battery or not communicating)**

RED **POWER** LED functions:

- Off: No AC power/No Back-up Battery Power
- Solid: AC power
- Flashing: **Running on Battery Back-up**

Programming Central Station Telephone Numbers & Account Codes:

The System must be programmed with the central station receiver telephone numbers and a 4 digit account code before unit is operational. The System can learn 2 telephone numbers: the alarm receiver's toll free number as the "primary number" and the alarm receiver's local long distance number as the "secondary number" as a back-up, in case the primary number is busy.

NOTES:

- Program the system using a touch-tone telephone connected on the same line as the Base unit. You will hear the Base unit speak each number that is pushed as you push them.
- **Dial slowly and listen for each number to be announced. If you did not hear the number announcement - the number was not recorded.**
- The System can learn numbers up to 32 digits long,
- If you chose to only program one central station number, simply enter "#" in place of the second number.
- If you need to add a pause in your number sequence, simply insert a "*" where ever you need a 1 second pause. Insert "***" if you need a 2 second pause.
- Remember to program a **1** before any long distance numbers.
- Remember to program any **AREA CODE** numbers if needed.
- Remember to program any **PREFIX NUMBERS** (like 9 to get an outside line) if needed.

Step 1: Write out the numbers you are wanting to program into the System. Below we will refer to the first phone number as the Primary number, the second phone number as the secondary number. Include area-codes and a 1 or 9 if they are needed for dialing – just like you would need to dial the customers phone from their house.

Step 2: Establish a phone connection with your cell phone. This is done as follows. From the home phone (on the same line as the System is hooked-up to) call your cell phone (or any other number - if you don't have a cellular telephone handy). Answer the phone (or wait until the number called answers the phone and tell them just to wait a minute while you program your System).

Note: If you don't have an open line when you begin to program the System with your telephone - the phone company will think you are trying to dial an outside number and will try to complete the call. By having the phone line already in use with your cell phone or other called party - you are now ready to program the System.

Step 3: With the phone connection established in Step 2, slide the slide switch at the back of the Base unit to **Programming Mode**. You will hear an audible announcement of this position.

Step 4: Start programming the System. Push the numbers slowly and listen to the voice prompts. Follow the next steps to enter the numbers.

NOTE: The “#” key (the key in the lower right corner of your phone keypad) must be entered after the passcode, after each central station number, and after the account code as shown in the steps below. **Do not ignore the “#” key.**

Central Station Number, Account Code, Caretaker Telephone Number Programming:

Note: You will be using your telephone keypad on an open telephone line to program the System numbers. The base will speak each number that is pushed on the telephones keypad. You DO NOT have to program a secondary (or back-up) central station telephone number, but we highly recommend it.

1. Call your cell phone **from the telephone on the same line as the Base unit**. Answer your cellphone and keep this line open for the following steps.
2. Slide Programming switch to Programming Mode
3. Enter (Passcode) – XXXX then “#” (Default is “1 2 3 4”)
[Voice prompt– Enter 1st phone number followed by “#”]
4. Enter primary central station number – X XXX XXX XXXX then “#”
[Voice prompt – Enter 2nd phone number followed by “#”]
5. Note: to skip secondary phone number - just enter “#” again
6. Enter secondary central station phone number
X XXX XXX XXXX then “#”
[Voice prompt – Enter 4 digit account number]
7. Enter 4 digit account number – AAAA then “#” (account number must be 4 digits long – no letters)
[Voice prompt – Enter Caretaker number]
8. Enter Caretaker Number – – X XXX XXX XXXX then “#”
[Voice prompt – Programming complete]
(This Caretaker number is the non-emergency number that the CaretakerSentry can call. This is generally programmed with a nurse or caregiver telephone number. To skip and not program a Caretaker number – just enter “#” again. If no number is programmed, the Caretaker button will work the same as the Emergency button)
8. Slide Programming Switch to Normal Operation (“Emergency Call Mode”)

To Change the 4 Digit Passcode:

The factory default for the programming passcode is “1 2 3 4”. If you wish to change this passcode to prevent unauthorized re-programming, simply follows these steps:

1. Establish a phone line as you did in Step 2 on page 7. Slide programming switch to Programming Mode.
2. Enter “1 2 3 4” or your old Passcode
XXXX then “* # * #”
[Voice prompt – Enter New Passcode]
3. Enter “Y Y Y Y” (your new passcode) then “#”
[Voice prompt – Programming complete]

- Slide Programming Switch to Normal Operation
[Voice prompt – “Emergency Call Mode”]

Note: there is no way to easily reset the passcode if the number is forgotten or mis-programmed, the unit will have to be exchanged. **A forgotten passcode IS NOT a warranty issue. Be careful when changing passcodes.**

Setting Programming Options:

As always - First establish a phone line as you did in Step 2 on page 8. Several options can be changed during one programming session but you must exit and re-enter the passcode each time. (You do not have to open a new phone line or slide the switch each time an option is set).

There are nine options that you may wish to set. These are as follows:

- **Option #0 – Voice Prompt Language Selection.** This changes the language of the voice prompts between **English, Spanish, French** (Canadian) and **Chinese**. **Factory setting is English.**
- **Option #1 -- Dial 9 Before Number.** This automatically dials 9 before any of the programmed numbers are dialed. Only use if 9 is used to dial an outside line. **Factory setting is NO.**
- **Option #2 -- Communication Protocol.** This is used to select between Contact ID or 4+2. Consult with your central station before changing. **Factory setting is Contact ID.**
- **Options Option #3 -- Dial *82 before the number.** This is used to temporarily disable Caller ID blocking. Consult with your central station before changing. **Factory setting is NO.**
- **Option #4 -- DEMO / Tradeshow Mode. THIS MODE IS FOR DEMO MODE ONLY FOR SALESPERSONS.** This feature is used to DEMO the unit without having access to a telephone line. Use this feature with a Viking DLE-200 test box. **NEVER LEAVE THE UNIT IN THIS MODE FOR NORMAL OPERATION. Factory setting is NO.**
- **Option #6 -- Automatic Call-Back Feature.** This feature allows the unit to call a central station, send the Emergency Event information, hang-up and then automatically answer a return call back. After the base connects with central station, transmits the Personal Emergency and account info and receives the Kiss-Off tone, the base hangs up. The CaretakerSentry will then answer the next incoming call and the Customer will speak with the central station with no indication that a callback has occurred. Consult with your central station before changing. **Factory setting is NO.**
- **Option #7 -- Send Restore Message.** This option automatically sends Restoral messages to the central station when a problem with the CaretakerSentry has been corrected. Consult with your central station before changing. **Factory setting is NO.**
- **Option #8 -- Periodic Test Interval.** This change the Periodic Test interval between Daily, Weekly and Monthly. Consult with your central station before changing. **Factory setting is Weekly.**
- **Option #9 -- Reset To Factory Settings.** This option erases all phone numbers, account numbers and reset all options back to the factory settings.

To Set the Options:

Option #0 – Voice Prompt Language – selects voice prompt language

1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
2. Enter – 0 [Voice prompt – Voice Prompt Language],
1 for English – FACTORY SETTING
2 for Spanish
3 for French
4 for Chinese
3. Enter either 1, 2 or 3
[Voice prompt 1 – English]

- [Voice prompt 2 – Spanish]
- [Voice prompt 3 – French]
- [Voice prompt 4 - Chinese]
- 4. Enter “#” (to exit programming mode)
[Voice prompt –Programming complete]
- 5. Slide Programming Switch to “Normal Operation”

Option #1 – Dial 9 Before Number - Dials '9' to get an outside line.

- 1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
- 2. Enter – 1
[Voice prompt – Dial 9 before number,
1 for YES – 3 for NO]
- 3. Enter either 1 or 3
[Voice prompt 1 – YES]
[Voice prompt 3 – NO] – FACTORY SETTING
- 4. Enter “#” (to exit programming mode)
[Voice prompt – Programming complete]
- 5. Slide Programming Switch to Normal Operation
[Voice prompt – “Emergency Call Mode”]

Option #2 – Alarm Communication Protocol – selects communication language with central station

- 1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
- 2. Enter – 2 [Voice prompt – Alarm Communication Protocol],
1 for CONTACT ID - FACTORY SETTING
2 for 4+2
- 3. Enter either 1 or 2
[Voice prompt 1 – CONTACT ID] [Voice prompt 2 – 4+2]
- 4. Enter “#” (to exit programming mode)
[Voice prompt – Programming complete]
- 5. Slide Programming Switch to “Normal Operation”
[Voice prompt – “Emergency Call Mode”]

Option #3 -- Dial *82 Before the Number - Displays unblocked Caller ID information from the caller for this call only.

- 1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
- 2. Enter – 3[Voice prompt – Dial *82 before number,
1 for YES – 3 for NO]
- 3. Enter either 1 or 3
[Voice prompt 1 – YES]
[Voice prompt 3 – NO] – FACTORY SETTING
- 4. Enter “#” (to exit programming mode)
[Voice prompt – Programming complete]
- 5. Slide Programming Switch to Normal Operation
[Voice prompt – “Emergency Call Mode”]

Option #4 – DEMO/Tradeshaw Mode – THIS MODE IS FOR DEMO MODE ONLY FOR SALESPERSONS.

This feature is used to DEMO the unit without having access to a telephone line. Use this feature with a Viking

DLE-200 test box. **NEVER LEAVE THE UNIT IN THIS MODE FOR NORMAL OPERATION.**

1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
2. Enter – 4
[Voice prompt – DEMO Tradeshow Mode],
1 for YES 3 for NO– FACTORY SETTING
3. Enter either 1 or 3
[Voice prompt 1 – YES] [Voice prompt 3 – NO]
4. Enter “#” (to exit programming mode)
[Voice prompt – Programming complete]
5. Slide Programming Switch to “Normal Operation”
[Voice prompt – “Emergency Call Mode”]

Option #6 -- Automatic Call-Back Feature - This feature allows the unit to call a central station, send the Emergency Event information, hang-up and then automatically answer a return call back. After the base connects with central station, transmits the Personal Emergency and account info and receives the Kiss-Off tone, the base hangs up but keeps the communication link active between the pendant/EWC and itself. It will then answer the next incoming call and the Customer will speak with the central station with no indication that a callback has occurred.

1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
2. Enter – 6
[Voice prompt – Automatic Call-back Mode],
1 for YES
3 for NO– FACTORY SETTING
3. Enter either 1 or 3
[Voice prompt 1 – YES] [Voice prompt 3 – NO]
4. Enter “#” (to exit programming mode)
[Voice prompt – Programming complete]
5. Slide Programming Switch to “Normal Operation”
[Voice prompt – “Emergency Call Mode”]

Option #7 – Send Restore Message - These are messages that are automatically sent to the central station to report the AC Power or the communication (RF or battery) problem with the Pendant has been restored or corrected.

1. Enter (Passcode) – XXXX then “**” (Default is **1234**) [Voice prompt – Option Mode]
2. Enter – 7
[Voice prompt – Send restore code to call center,
1 for YES – 3 for NO]
3. Enter either 1 or 3
[Voice prompt 1 – YES]
[Voice prompt 2 – NO] – FACTORY SETTING
4. Enter # (to exit programming mode)
[Voice prompt – Programming complete]
5. Slide Programming Switch to Normal Operation
[Voice prompt – “Emergency Call Mode”]

Option #8 - Periodic Test Interval (for automatic check-in with central station)

1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
2. Enter – 8

- [Voice prompt – Periodic test interval,
1 for 1 day,
2 for 7 days - FACTORY SETTING
3 for 30 days]
- 3. Enter 1, 2 or 3
[Voice prompt 1 – Every day] [Voice prompt 2 – Every 7 days] [Voice prompt 3 – Every 30 days]
- 4. Enter “#” (to exit programming mode)
[Voice prompt – Programming complete]
- 5. Slide Programming Switch to “Normal Operation”
[Voice prompt – “Emergency Call Mode”]

Option #9 – Reset To Factory Settings

- 1. Enter (Passcode) – XXXX then “**” (Default is **1234**)
[Voice prompt – Option Mode]
- 2. Enter 9
[Voice prompt – Reset to factory settings,
1 for YES – 3 for NO]
- 3. Enter either 1 or 3
[Voice prompt 1 – YES] [Voice prompt 3 – NO]
- 4. Enter “#” (to exit programming mode)
[Voice prompt – Programming complete]
- 5. Slide Programming Switch to “Normal Operation”
[Voice prompt – “Emergency Call Mode”]

Note: The telephone numbers are erased, so if the Emergency button is pushed, you will hear the dial tone and then “Hanging Up”

REMOTELY PROGRAMMING THE System:

The System can be remotely programmed over a telephone line in one of two ways – Standard and Unaided Remote Programming.

Standard Method:

- 1. Simply call the number of the residence where the System is connected.
- 2. Have the user slide the switch at the back of the base unit to “Programming Mode”.
- 3. Then, proceed to program the System as needed with the touch-tones of the phone on which you are calling from.
- 4. Have them slide the switch back to Normal Operation when done.

Unaided Remote Programming:

- 1. The dealer calls the customers phone number that the CaretakerSentry is installed on.
- 2. The dealer allows the phone to ring 1 ring cycle - (equivalent to 1 ring back tone - about 6 seconds of ringing).
- 3. The dealer then hangs up.
- 4. The dealer waits 10 seconds and dials the customer phone number again. The CaretakerSentry sees this second call within a 30 second period and immediately answers the phon.
 - The CaretakerSentry will only start this 30 second timer - waiting for a second call to ring the phone ONLY if the phone only rings 1 time. If the phone rings more than 1 time, the CaretakerSentry unit will not start the timer and will NOT wait for the second call - feature is canceled and reset.
- 5. As soon as the CaretakerSentry answers the call, [voice prompt – programming mode] is played.
- 6. The dealer enters the 4 digit passcode and then CaretakerSentry enters programming mode (unit will program normally).

7. If no passcode or an incorrect passcode is entered within 15 seconds, the CaretakerSentry unit will hang up and reset.
8. The dealer will press '9' to end programming, exit and reset the CaretakerSentry unit.

Note: The programming voice prompts will still be heard from the base while the unit is being remotely programmed – this is designed to allow the customer know that changes are being made to their unit.

Adding Additional Pendants to the System:

Note:

When replacing pendants – due to loss or malfunction, **ERASE ALL PENDANTS** and re-learn just the pendant that will be used. All pendants are **supervised** and the base will continue to “look” and send failure messages to the central station for pendants that are no longer working or not currently being used. **SEE SECTION – ERASING ALL PREVIOUSLY LEARNED PENDANTS**

The System comes with one Pendant. If you want to add additional pendants, follow the directions here. If not, skip this section.

The Pendant's unique ID code should already be learned into the Base unit when it came from the factory. You can add additional Pendants (up to 5 Two-way pendants and Emergency Wall Communicators plus 4 Basic pendants, for a total of 9) to the System by simply teaching each Pendant's ID code to the Base unit.

When more than 1 Pendant is learned the base, the base unit will announce the Pendant number - i.e.” Pendant #X, battery is OK” – when only 1 pendant is learned, the base will only announce “Battery is OK”

Anytime the Pendant reports to the call center (even with pendant battery status), the pendant number information is also transmitted to the central station – Zone 1, 2, 3,... for Pendants 1, 2, 3,...

To program a new Pendant to the Base unit:

On the Base Unit:

1. Push and **release** the RED Code Learn button on the back of the Base unit. You will hear a voice prompt “Pendant Learning”



On the Pendant:

2. PUSH the Blue Emergency Button and the Gray Test Button on the Pendant at the same time and hold until you hear “Pendant learning” again.
 - If “Base and Pendant out of Range” or, “Pendant Learning Failed” is heard from the pendant, STOP - wait 30 seconds and start at step #1 again.

3. When the Pendant Learning has been successful you will hear “Pendant Code Learned”
 - If “Base and Pendant out of Range” or, “Pendant Learning Failed” is heard from the Base or Pendant, STOP - wait 30 seconds and start at step #1 again

Notes:

- You have less than 45 seconds after pushing the RED Code Learn button on the base unit to get the pendant to enter “Pendant Learning” mode.
- Teaching additional pendants does not erase previously learned pendants. It adds it to the list. The base unit will automatically exit the pendant learning mode after several seconds or if Pendant Learning Failed
- If you hear the voice prompt “Base and Pendant Out of Range” or, “Pendant Learning Failed” an error has occurred, wait 30 seconds for the base to reset and repeat steps 1 & 2.
- Only 9 Basic Pendants, or combined with up to 5 Two-Way and/or EWCs for a total of 9, can be taught to a base (If Pendants have been lost or replaced, we recommend you erase all pendants and re-learn all current Pendants.)

Erasing All Previously Learned Pendants:

Simply holding down the **RED Code Learn** button at the back of the Base unit for 8 seconds. You will hear the confirmation of "Erase All Pendants, All pendants Erased" from the Base unit. After that, you can program a new pendant to the base unit following the steps above as desired.

Basic Set-up and Test:

Lanyard, Wrist Strap/Beltclip

The System comes with 3 means of carrying the pendant. You can wear the Pendant using the Lanyard, Wrist Strap. or Beltclip. Note that the Wrist Strap can also be used as a way to attach the Pendant to a wheel chair. Choose the method that suits your customer's needs.

Note that the lanyard has a break-away feature to release if it gets caught on something while around someone's neck, and that the holster for the Beltclip and Wrist Strap are one piece. The holster screws onto either the clip or the strap.

System Check:

The button on the back of the pendant case which is used to perform the BATTERY CHECK, will also perform a SYSTEM CHECK. Press and hold this button for more than four (4) seconds. One of the following announcements will be made.

"All systems are ok."

This confirms the following:

Battery status.

The Pendant is working.

The Base unit is working.

The Base unit is connected to a working phone line.

The Pendant is within range of the Base unit.

"System cannot detect dial tone" / "System cannot detect phone line."

The Base unit is not connected to the phone line or can not get a dial tone for some reason.

"Base and Pendant out of range."

This confirms one or more of the following:

The Pendant has malfunctioned.

The Base has malfunctioned.

The Pendant and Base unit are not communicating with each other or are out of range.

If you receive a failure notification, please refer to the TROUBLESHOOTING guide on page 17 of this manual to determine the cause.

Remind your customers to perform a System Check at least twice a month.

NOTE:
The Pendant does a self test on the pendant battery every 13 hours. When the battery level gets to the "Replace battery now" level, it will send a signal to the central station to notify them of this situation.

This testing feature is very useful - allowing you to confirm that the system has coverage from all parts of the home. Simply walk to the various parts of your home and push the Battery Check /

System Check button until you hear the dial tone. You can then release the button, and you will hear "System Check: All Functions are OK" voice prompt, or "Pendant and Base out of Range"

The Base also counts these 13 hour checks. If 4 reports are missed in a row (52 hours later), the Base unit will send a loss of RF signal to the central station.

NOTE: If your telephone service offers voice mail and the voice mail System uses a Stutter Dial Tone to notify you of message, the System test may fail – "cannot detect dial tone" if a message is waiting. This is normal as the CaretakerSentry is "listening" for a conventional dial tone. Check your messages so the dial tone returns to a solid tone and test again. In the event of an emergency, the CaretakerSentry will dial the emergency number regardless of a standard or stutter dial tone.

Operating Range:

The operating range of the System will depend on many factors including intervening walls, electrical interference or various appliances such as vacuum cleaners, refrigerators, microwave ovens, mixers, coffee grinders, hair dryers and other sources of electrical noise around the house. The System should cover the entire home and a short distance into the front, side and back yards. Perform the System Check on page 15 to determine the boundaries of the System to know the limits of operation in your home and surroundings.

Testing:

Be sure to have your subscriber (customer) fully test the System with the central station. **Please consult with your customer and central station on procedures and policies if the central station receives a emergency call but they cannot communicate with your customer.**

Out of Service Storage or Transportation:

When storing the Base unit -or taking it on a trip - or sending it in for repair - it is recommended that you turn the Master ON / OFF switch to OFF so that the "Running on battery power" announcement doesn't repeat for the next 24 hours. Remember to turn the Master ON / OFF switch ON in the Base unit when it is put back in service.

Caution - If the Line Seizure feature is used and installed the installer, customer or user must verify that the telephone line and equipment will work properly when the System is turned OFF and/ or removed from service.

TROUBLESHOOTING GUIDE:

What if I can't get the System to work?

Step 1 – Push the Black RESET button with a pen at the back of the Base unit (next to the power cord). Check the System. If it is now working skip the remaining steps.

Step 2 – Check the PENDANT.

What to do:

Press the Gray Battery Test / System Check button on the back of the Pendant.

What to expect:

A voice should announce the condition of the battery.

A light on the front of the Pendant will come on.

If nothing happens, the most likely cause is that the battery is dead or has been installed backwards.

If the Pendant says "Battery ok", go on to step 3.

Step 3 – Check the connections on the BASE UNIT

What to do:

Make sure the AC Adapter is plugged into a live AC outlet and into the Base unit.

Make sure the Master ON / OFF switch is ON.

What to expect:

The Red POWER light will be on if the Base unit is connected to power.

If the POWER light is on and the Base unit is connected to a working phone line, go on to step 3.

Step 4 – Make sure the Pendant and the Base unit are communicating with each other.

Background information: The Pendant and the Base unit are connected by a wireless radio link. They must be synchronized to work together. They should have already been synchronized at the factory, but resetting this is simple. Step 4 – Make sure the Pendant and the Base unit are communicating with each other.

Background information: The Pendant and the Base unit are connected by a wireless radio link. They must be synchronized to work together. They should have already been synchronized at the factory, but resetting this is simple.

What to do: Follow the steps on page 14: Adding Additional Pendants to the System

Now perform the System Check test again (page 15) to make sure all is well.

If the System is still not working, call tech support. See page 24 for telephone number.

Central Station Items of Interest:

Number of call attempts: 9 times - alternating between the first and second numbers programmed.

Supervision of Pendants Yes – all pendants check in with the Base every 13 hours. A Pendant missing for more than 52 hours signals a “Loss of RF” condition/report.

Reporting of Low Pendant Battery to Central Station
Upon receipt of first “Replace Battery Now” Reporting of AC Power Loss After 8 hours of power loss - and only once. Reporting of Restore Codes Only once after restore condition met.
(Dealer programmable feature).

Operator Commands The CareTakerSentry has the ability to switch between **Simplex** (CB radio like) and **Duplex** (Telephone like) modes - **The default setting is Duplex.** The central station operator can switch to Simplex when they are having difficulty hearing or being understood. When in Duplex mode Operator will talk with the customer just as they would during a normal telephone call.
When in Simplex mode – the operator will need to toggle:
1 = Talk
3 = Listen
4 = Switches CaretakerSentry back to Duplex
7 = Extend VOX timer
9 = End the call (unit will hang up and reset)

Contact ID Event Code Profile: Standard Ademco Contact ID

Contact ID Event Codes Sent:

- 1-101 Personal Emergency (personal emergency)
- 1-606 Listen-in to follow (request for voice call)
- 1-301 AC Loss (AC power loss longer than 8 hours)
- 3-301 Restoral - AC Loss
- 1-381 Loss of Radio supervision (pendant has not reported to the base in the last 52 hours, four 13 hour cycles)
- 3-381 Restoral - Loss of Radio supervision
- 1-384 RF low battery (pendant batter is < 3.5 volts)
- 3-384 Restoral - RF low battery
- 1-602 Periodic test report (periodic intervals are installer programmable to 1, 7 or 30 days - default is 7 days)

Note: Each pendant is set up as a Zone - if a pendant event code is sent, the central station can know which pendant initiated the call (only if the installer notifies the central station which person carries each pendant).

4+2 Event Codes Profile:

- 01, 02, 03, 04...09 Medical Alarms, zones 1, 2, 3, 4...9
- 40 A/C Fail
- 50 A/C Restoral
- 61, 62, 63, 64...69 R/F Loss of Supervision, zones 1, 2, 3, 4...9
- 71, 72, 73, 74...79 R/F Restoral of Supervision, zones 1, 2, 3, 4...9
- 41, 42, 43, 44...49 R/F Low Battery, zones 1, 2, 3, 4...9
- 51, 52, 53, 54...59 R/F Battery Restoral, zones 1, 2, 3, 4...9
- 30 Timer Test

Note: Each pendant & EWC is treated as a separate zone

Information The FCC Wants You To Know:

CaretakerSentry Base Unit - Model 40914

FCC ID: TYD-CS40914

IC ID: 8471A-CS40914

CaretakerSentry Two-Way Pendant - Model 41915

FCC ID: TYD-CS41915

IC ID: 8471A-CS41915

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

This portable transmitter with its antenna complies with FCC/IC RF exposure limits for general population/uncontrolled exposure.

This Device complies with Industry Canada License-exempt RSS standard(s). Operation is subject to the following two conditions: 1) this device may not cause interference, and 2) this device must accept any interference, including interference that may cause undesired operation of the device.

Cet appareil est conforme à Industrie Canada exemptes de licence RSS ou les normes. Opération est sujette aux deux conditions suivantes: 1) cet appareil ne doit pas brouillage et 2) cet appareil doit accepter toute interférence reçue, y compris les interférences pouvant entraîner un fonctionnement indésirable du dispositif.

This equipment complies with Part 68 of the FCC rules and the requirements adopted by the ACTA:

US: TYD-CS40914

REN: 3.7

Privacy of communications may not be ensured when using this product.

This product meets the applicable Industry Canada technical specifications. / Le présent matériel est conforme aux spécifications techniques applicables d'Industrie Canada.

The Ringer Equivalence Number (REN) is an indication of the maximum number of devices allowed to be connected to a telephone interface. The termination of an interface may consist of any combination of devices subject only to the requirement that the sum of the RENs of all the devices not exceed five. / L'indice d'équivalence de la sonnerie (IES) sert à indiquer le nombre maximal de terminaux qui peuvent être raccordés à une interface téléphonique. La terminaison d'une interface peut consister en une combinaison quelconque de dispositifs, à la seule condition que la somme d'indices d'équivalence de la sonnerie de tous les dispositifs n'excède pas cinq.

General Warnings And Precautions!

Warning!

- LogicMark, LLC DOES NOT represent this unit to be waterproof. To reduce the risk of damage to the unit, DO NOT expose the pendant or base unit to prolonged water, rain or moisture.
- This System is intended for residential use only. It may not work properly if connected to some commercial phone Systems.
- This System works with tone dialing Systems only. Do not use if you have pulse dialing only. The System cannot be connected to cellular telephone service, MagicJack, Skype or similar virtual telephone services.
- This product and/or accessories may contain lead, a chemical known to the State of California to cause birth defects or other reproductive harm.

Note: The System does not interfere with pacemakers.

Important Safety Instructions:

When using your System , please observe the following instructions in order to reduce the risk of personal injury, electrical shock, or fire.

1. Save these instructions for future reference.
2. Read all instructions carefully and make sure you understand them.
3. Unplug this product from the electrical outlet before attempting to clean it. Do not use any liquids for cleaning either the Base unit or the Pendant. Use only a soft damp cloth.
4. Do not locate the Base unit near any source of water, such as in the bathroom, near a sink in the kitchen, or the like.
5. Do not open the unit, (except the battery compartments as directed in these instructions) on the Pendant for any reason. Any attempt to open the case of the Base unit or the Pendant will void the warranty and may damage the unit or prevent it from working properly.
6. Do not use any power source other than that supplied with this product specifically for this product. Use of any other power source could damage your unit or make it unsafe.
7. Always dispose of batteries properly. Never throw batteries into a fire. Consult your local ordinances for proper battery safety.
8. If your unit becomes damaged do not attempt to repair it. Please return it to your dealer or service provider for service or replacement as stipulated on the next page under "ONE YEAR LIMITED WARRANTY".

LogicMark Disclaimers and Limited Warranty

COMMUNICATION AND RESPONSE LIMITATIONS: Purchaser acknowledges that signals which are transmitted over telephone lines, or other modes of communication pass through communication networks wholly beyond the control of LogicMark LLC and are not maintained by LogicMark LLC, and, therefore, LogicMark LLC shall not be responsible for any equipment or communication failure which prevents transmission signals from reaching your contact list including emergency 911 operators or damages arising therefrom.

Purchaser acknowledges that LogicMark LLC provides no response to the System's equipment. The equipment is designed to communicate with the a central station of your choice and LogicMark LLC is not and shall not be responsible for ambulance, police or other emergency response time or that any response will be provided by the central monitoring station.

ELECTRIC AND TELEPHONE SERVICE IS NECESSARY AND PURCHASER'S RESPONSIBILITY: Purchaser acknowledges that the System's equipment plugs into a standard land line telephone jack and communicates over standard telephone lines using two way voice communication. Purchaser acknowledges that the System's equipment requires 110 Volt AC power and electrical outlets and receptacles, telephone hook-ups, RJ11 Block or equivalent, all of which is Purchaser's responsibility to obtain and maintain.

TESTING AND SERVICE OF THIS EQUIPMENT: The equipment, once installed, are in the exclusive possession and control of the Purchaser, and it is Purchaser's sole responsibility to test the operation of equipment and request warranty service if the equipment is under warranty.

PURCHASER'S EXCLUSIVE REMEDY: Purchaser's exclusive remedy for LogicMark LLC's default hereunder is to require LogicMark LLC to repair or replace, at LogicMark LLC's option, any equipment or part of the personal emergency alert System which is non-operational during LogicMark LLC's warranty period.

LIMITATION OF LIABILITY: This equipment is not designed or guaranteed to prevent any loss or injury. This Limited Warranty and Disclaimer of Liability constitutes the terms of sale and use of the System (and accessory) equipment, and if, notwithstanding these terms of sale and use of the product, there should arise any liability on the part of LogicMark LLC as a result of any cause whatsoever, regardless of whether or not such loss, damage, or personal injury was caused by or contributed to by LogicMark LLC's negligence to any degree or failure to perform any obligation or strict products liability, such liability will be limited to an amount paid by the Purchaser to LogicMark LLC for the product, or to the sum of \$350.00, whichever is greater.

ONE YEAR LIMITED WARRANTY

Warranty: LogicMark, LLC warrants to the original consumer/purchaser that this product shall be free of defects in material and workmanship under normal use and circumstances for a period of **one** (1) year from the date of original purchase for use.

How to Get Warranty Service: All warranty service must go through your dealer or service provider.

Warranty Limitations and Exclusions: LogicMark LLC does not represent nor warrant that this System will prevent any loss, damage or injury to person or property, or that the personal emergency alert System will in all cases provide the protection for which it is installed or intended. Purchaser acknowledges that LogicMark LLC is not an insurer, and that Purchaser assumes all risk for loss or injury to Purchaser's property or person. LogicMark LLC has made no representation or warranties, except those expressed herein and hereby disclaims any express warranty of merchantability or fitness for any particular use.

LogicMark, LLC. will not be responsible for the improper use of this System, nor will it be responsible for failure resulting from the use of other equipment connected to the same phone line. We will not be responsible for the quality of the phone line or the reliability or quality of the phone service with which the System is used. LogicMark, LLC. will not be responsible for the installation of the System. It will not be responsible for the improper use or abuse of the base unit or pendant.

This warranty shall constitute the sole liability of LogicMark, LLC concerning the product. ALL IMPLIED WARRANTIES INCLUDING, WITHOUT LIMITATION, THE WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE LIMITED IN DURATION TO THE TERM OF THIS EXPRESSED ONE YEAR LIMITED WARRANTY. NO PERSON, FIRM, OR CORPORATION IS AUTHORIZED TO ASSUME FOR LogicMark, LLC. ANY OTHER LIABILITY IN CONNECTION WITH THE SALE AND USE OF THE PRODUCT. LogicMark, LLC AND AGENTS AND DISTRIBUTORS WILL BEAR NO LIABILITY WHATSOEVER FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES OR CHARGES OF ANY KIND

This warranty is void if the product has been damaged or tampered with or if the product or any such parts have been opened. In all cases of damage during shipment, a claim must be filed with the shipping carrier and not with LogicMark, LLC.

State Law: This warranty gives you specific legal rights; you may also have other rights which vary from state to state. Some states do not allow the exclusion or limitation of incidental or consequential damages, or a limitation on the duration of implied warranties, so the above disclaimers may not apply to you

OUT OF WARRANTY REPAIRS

As with all warranty repairs, you need to work through your dealer or service provider. If the warranty period has expired or if you are not the original owner of the product, you can return the product pre-paid to LogicMark, LLC; 10106 Bluegrass Pkwy; Louisville KY, 40299, USA for repair or replacement at our standard rates.

Replacement Parts / Optional Accessories



Emergency Wall Communicator
Part #41920



Additional Two-Way Pendants
Part #41915



Additional Waterproof Pendants
Part #40915



Rechargeable Li-Ion Batteries - 2 Pieces
Part #35917



Lock Box
Part #30913



Rechargeable AA Batteries
4 Pieces
Part #35918



Cellular Accessory
Part #40914ACC

Product Specifications:

RF Characteristics:	DECT 6.0 Two-way voice pendant
Communication Protocols	Two options: Ademco Contact ID / 4+2
Operating Range	Covers your typical American house and into the front, back, and side yards. Up to 600 feet line-of-sight from Base unit.
Power Adapter Ratings	Input: 110 VAC Output: 7.5 volts DC - 800 mA
Base Unit Power Consumption	60 mA in standby mode 200 mA when dialing
Back-up Battery Supply	AA NiMH 2,400 mAh batteries (4 pieces). Should be replaced every several years to maintain proper charge during standby times. Back-up Operation Duration: 24 hours with fresh batteries
Dialing Style	Touch-Tone only
Telephone Number Length	32 digits maximum
Telephone Numbers	Can learn up to 2 central station numbers, both a Primary and a Secondary back-up.
Pendant Battery Life	4-6 months between charges (estimated)
Operating Temperature	32° to 120° F. (0° to 49° C).
Pendant water-resistance	Water Spray Standard IP-X5 (Water spray test).
Number of learned Pendants per Base pendants for a total of 9 pendants	9 Basic pendants or up to 5 Two-way pendants and/or EWCs and 4 Basic



10106 Bluegrass Pkwy
Louisville KY, 40299
Tel: 1-502-410-0402
Toll Free: 1-888-283-6516
Fax: 1-703-934-7935
www.LogicMark.com